This is a free page sample. Access the full version online.

AS 60118.8—2007
IEC 60118-8, Ed. 2.0:2005
Reconfirmed 2019

Australian Standard®

Hearing aids

Part 8: Methods of measurement of performance characteristics of hearing aids under simulated in situ working conditions

STANDARDS
Australia

This is a free page sample. Access the full version online.

This Australian Standard® was prepared by Committee AV-003, Acoustics—Human Effects. It was approved on behalf of the Council of Standards Australia on 3 November 2006. This Standard was published on 20 February 2007.

The following are represented on Committee AV-003:

- Association of Accredited Certification Bodies
- Association of Australian Acoustical Consultants
- Association of Consulting Engineers Australia
- Audiological Society of Australia
- Australasian Faculty of Occupational Medicine
- Australian Acoustical Society
- Australian Chamber of Commerce and Industry
- Australian Council of Trade Unions
- Australian Industry Group
- Department of Consumer & Employment Protection, WorkSafe Division, WA
- Department of Labour, NZ
- National Acoustic Laboratories
- New South Wales Nurses, Association
- New Zealand Audiological Society
- Royal Institute of Naval Architects
- Safety Institute of Australia
- The Australian Society of Otolaryngological Head and Neck Surgery
- Victorian WorkCover Authority
- WorkCover New South Wales

This Standard was issued in draft form for comment as DR 06473.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through public comment period.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting www.standards.org.au

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

This is a free page sample. Access the full version online.

STANDARDS AUSTRALIA

RECONFIRMATION

OF

AS 60118.8—2007

Hearing aids

Part 8: Methods of measurement of performance characteristics of hearing aids under simulated in situ working conditions

RECONFIRMATION NOTICE

Technical Committee AV-003 has reviewed the content of this publication and in accordance with Standards Australia procedures for reconfirmation, it has been determined that the publication is still valid and does not require change.

Certain documents referenced in the publication may have been amended since the original date of publication. Users are advised to ensure that they are using the latest versions of such documents as appropriate, unless advised otherwise in this Reconfirmation Notice.

Approved for reconfirmation in accordance with Standards Australia procedures for reconfirmation on 22 August 2019.

The following are represented on Technical Committee AV-003:

Accident Compensation Corporation (New Zealand)
Association of Australasian Acoustical Consultants
Australian Acoustical Society
Australian Chamber of Commerce and Industry
Australian Council of Trade Unions
Australian Hearing
Department of Defence (Australian Government)
Engineers Australia
Ministry of Health (NZ)
New South Wales Nurses Association
New Zealand Audiological Society
Worksafe Division, Department of Commerce, Western Australia
WorkSafe Victoria

This is a free page sample. Access the full version online.

NOTES

This is a free page sample. Access the full version online.

Australian Standard®

Hearing aids

Part 8: Methods of measurement of performance characteristics of hearing aids under simulated in situ working conditions

Originated as part of AS 1088—1971.
Previous edition AS 1088.8—1987.
Revised and redesignated as AS 60118.8—2007.

COPYRIGHT

© Standards Australia

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 8036 9

SAI GLOBAL

This is a free preview. Purchase the entire publication at the link below:

Product Page

Looking for additional Standards? Visit Intertek Inform Infostore

Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation