



National Standards Authority of Ireland

IRISH STANDARD

I.S. CEN/TS 15360:2006

ICS 65.080

**FERTILIZERS - DETERMINATION OF
DICYANDIAMIDE - METHOD USING
HIGHPERFORMANCE LIQUID
CHROMATOGRAPHY (HPLC)**

National Standards
Authority of Ireland
Glasnevin, Dublin 9
Ireland

Tel: +353 1 807 3800
Fax: +353 1 807 3838
<http://www.nsai.ie>

Sales
<http://www.standards.ie>

*This Irish Standard was
published under the
authority of the National
Standards Authority of
Ireland and comes into
effect on:
May 15, 2006*

**NO COPYING WITHOUT NSAI
PERMISSION EXCEPT AS
PERMITTED BY COPYRIGHT
LAW**

© NSAI 2006

Price Code E

Údarás um Chaighdeán Náisiúnta na hÉireann

TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

CEN/TS 15360

March 2006

ICS 65.080

English Version

Fertilizers - Determination of dicyandiamide - Method using high-performance liquid chromatography (HPLC)

Engrais - Détermination de la teneur en dicyandiamide -
Méthode par chromatographie liquide à haute performance
(HPLC)

Düngemittel - Bestimmung von Dicyandiamid - Verfahren
mit Hochleistungs-Flüssigchromatographie (HPLC)

This Technical Specification (CEN/TS) was approved by CEN on 30 January 2006 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

CEN/TS 15360:2006 (E)

Contents	Page
Foreword.....	3
1 Scope	4
2 Normative references	4
3 Principle	4
4 Reagents	4
5 Apparatus	4
6 Sampling and sample preparation	5
7 Procedure	5
8 Calculation.....	5
9 Precision	6
10 Notes on procedure	7
11 Test report	7
Annex A (informative) Results of the inter-laboratory trial	8
Bibliography	9

Foreword

This Technical Specification (CEN/TS 15360:2006) has been prepared by Technical Committee CEN/TC 260 "Fertilizers and liming materials", the secretariat of which is held by DIN.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CEN/TS 15360:2006 (E)

1 Scope

This Technical Specification specifies a method for the selective determination of dicyandiamide (DCD) in addition to all the other forms of nitrogen fixations, particularly in fertilizers to which DCD has been added as a nitrification inhibiting agent.

2 Normative references

The following referenced documents are indispensable for the application of this Technical Specification. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

prEN 1482-2, *Fertilizers and liming materials — Sampling and sample preparation — Part 2: Sample preparation*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696:1987)*

3 Principle

The sample is dissolved or suspended in water, using an ultrasonic bath. To the filtered solution, methyl dicyandiamide is added as the internal standard, the solution is transferred onto a C18 reversed-phase column using a bypass injector and then separated. For the detection an UV-detector is used at a wavelength of 220 nm.

4 Reagents

Use only reagents of recognized analytical grade, and water conforming to grade 2 of EN ISO 3696.

4.1 Dicyandiamide standard solution

Weigh 50 mg of dicyandiamide of known purity into a 1 000 ml volumetric flask and dissolve in water, and make up the volume to the mark. Pipette 10 ml of this solution into a 100 ml volumetric flask and, after having added 10 ml of internal standard solution (4.2), make up the volume to the mark with water.

4.2 Internal standard solution

Weigh 50 mg of methyl dicyandiamide of known purity into a 1 000 ml volumetric flask, dissolve in water and make up the volume to the mark.

4.3 Methanol

HPLC-grade purity

5 Apparatus

5.1 Ultrasonic bath

5.2 Membrane filter

0,45 µm, with the usual filtration equipment.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-